



# **ELISA Kit for Chemokine C-X-C-Motif Receptor 3** (CXCR3)

Catalog No: tcue176



### **Available Sizes**

Size: 96T



# **Specifications**

#### **Research Area:**

Signal transduction; CD & Adhesion molecule; Tumor immunity; Infection immunity; Immune molecule;

#### **Species Reactivity:**

Homo sapiens (Human)

#### **Sample Type:**

Serum, plasma, tissue homogenates and other biological fluids

# **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.062ng/mL

# **Detection Range:**

0.156-10ng/mL

#### **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

## **Tested Application:**

ELISA

#### **SwissProt:**


P49682





# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Chemokine C-X-C-Motif Receptor 3 (CXCR3). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Chemokine C-X-C-Motif Receptor 3 (CXCR3). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Chemokine C-X-C-Motif Receptor 3 (CXCR3), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Chemokine C-X-C-Motif Receptor 3 (CXCR3) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!