

ELISA Kit for Myeloid Differentiation Factor 88 (MyD88)

Catalog No: tcue976

Available Sizes

Size: 96T



Research Area:

Signal transduction; CD & Adhesion molecule; Tumor immunity; Pulmonology;

Species Reactivity:

Homo sapiens (Human)

Sample Type:

Tissue homogenates, cell lysates and other biological fluids.

Sensitivity:

The minimum detectable dose of this kit is typically less than 0.055ng/mL

Detection Range:

0.156-10ng/mL

Assay Time: 3h

Detection Method:

Enzyme-linked immunosorbent assay for Antigen Detection.

Tested Application:

ELISA


SwissProt: Q99836

Copyright 2021 Taiclone Biotech Corp.



Test Principle

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Myeloid Differentiation Factor 88 (MyD88). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Myeloid Differentiation Factor 88 (MyD88). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Myeloid Differentiation Factor 88 (MyD88), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Myeloid Differentiation Factor 88 (MyD88) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!

Copyright 2021 Taiclone Biotech Corp.