

# ELISA Kit for Carbohydrate Antigen 19-9 (CA19-9)

Catalog No: tcue1262



## Available Sizes

**Size:** 96T



## Specifications

### Research Area:

Tumor immunity;Hepatology;Gastroenterology;

### Species Reactivity:

Homo sapiens (Human)

### Sample Type:

Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.

### Sensitivity:

The minimum detectable dose of this kit is typically less than 0.28U/mL

### Detection Range:

0.78-50U/mL

### Assay Time:

3h

### Detection Method:

Enzyme-linked immunosorbent assay for Antigen Detection.

### Tested Application:

ELISA


### SwissProt:

P21217

## Test Principle

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated

with an antibody specific to Carbohydrate Antigen 19-9 (CA19-9). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Carbohydrate Antigen 19-9 (CA19-9). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Carbohydrate Antigen 19-9 (CA19-9), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450\text{nm} \pm 10\text{nm}$ . The concentration of Carbohydrate Antigen 19-9 (CA19-9) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!