



ELISA Kit for Vitamin D Receptor (VDR)

Catalog No: tcue1358



Available Sizes

Size: 96T



Specifications

Research Area:

Metabolic pathway; Endocrinology;

Species Reactivity:

Homo sapiens (Human)

Sample Type:

tissue homogenates, cell lysates and other biological fluids

Sensitivity:

The minimum detectable dose of this kit is typically less than 0.251ng/mL

Detection Range:

0.625-40ng/mL

Assay Time:

3h

Detection Method:

Enzyme-linked immunosorbent assay for Antigen Detection.

Tested Application:

ELISA

SwissProt:

P11473


Test Principle

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated





with an antibody specific to Vitamin D Receptor (VDR). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Vitamin D Receptor (VDR). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Vitamin D Receptor (VDR), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of $450 \text{nm} \pm 10 \text{nm}$. The concentration of Vitamin D Receptor (VDR) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!