

# ELISA Kit for Fibroblast Growth Factor 9 (FGF9) Catalog No: tcue1518

**Available Sizes** 

**Size:** 96T

Specifications

#### **Research Area:**

Cytokine;Tumor immunity;Infection immunity;

**Species Reactivity:** Mus musculus (Mouse)

#### Sample Type:

serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

#### Sensitivity:

The minimum detectable dose of this kit is typically less than 6.2pg/mL

#### **Detection Range:**

15.6-1,000pg/mL

#### Assay Time:

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

### **Tested Application:**

ELISA

#### SwissProt:


P54130

## **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated



with an antibody specific to Fibroblast Growth Factor 9 (FGF9). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Fibroblast Growth Factor 9 (FGF9). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Fibroblast Growth Factor 9 (FGF9), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Fibroblast Growth Factor 9 (FGF9) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!

Copyright 2021 Taiclone Biotech Corp.