



# **ELISA Kit for Immunoglobulin E (IgE)**

Catalog No: tcue1641

	-
J	τ

#### **Available Sizes**

Size: 96T



# **Specifications**

#### **Research Area:**

Infection immunity; Immune molecule; Hematology; Autoimmunity;

## **Species Reactivity:**

Mus musculus (Mouse)

# **Sample Type:**

serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids

## **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.37ng/mL

# **Detection Range:**

0.78-50ng/mL

#### **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

## **Tested Application:**

**ELISA** 

## **SwissProt:**

P06336


# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated





with an antibody specific to Immunoglobulin E (IgE). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Immunoglobulin E (IgE). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Immunoglobulin E (IgE), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450 \text{nm} \pm 10 \text{nm}$ . The concentration of Immunoglobulin E (IgE) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!