



## **ELISA Kit for Nicotinamide Adenine Dinucleotide Phosphate Oxidase 1 (NOX1)**

Catalog No: tcue1657

Q9WV87


| Available Sizes   |
|---|
| Size: 96T   |
| Specifications  |
| Research Area: Enzyme & Kinase;Neuro science;Developmental science;                   |
| Species Reactivity:<br>Rattus norvegicus (Rat)  |
| Sample Type:<br>tissue homogenates, cell lysates and other biological fluids          |
| Sensitivity: The minimum detectable dose of this kit is typically less than 0.56ng/mL |
| Detection Range:<br>1.56-100ng/mL   |
| Assay Time:<br>3h   |
| <b>Detection Method:</b> Enzyme-linked immunosorbent assay for Antigen Detection.     |
| Tested Application: ELISA   |
| SwissProt:  |





## **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Nicotinamide Adenine Dinucleotide Phosphate Oxidase 1 (NOX1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Nicotinamide Adenine Dinucleotide Phosphate Oxidase 1 (NOX1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Nicotinamide Adenine Dinucleotide Phosphate Oxidase 1 (NOX1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of Nicotinamide Adenine Dinucleotide Phosphate Oxidase 1 (NOX1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!