



# **ELISA Kit for Interleukin 7 Receptor (IL7R)**

Catalog No: tcue1761

	J	U	
7			

#### **Available Sizes**

Size: 96T



# **Specifications**

#### **Research Area:**

Signal transduction; CD & Adhesion molecule; Infection immunity; Immunodeficiency;

## **Species Reactivity:**

Rattus norvegicus (Rat)

### **Sample Type:**

Serum, plasma and other biological fluids.

## **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.116ng/mL

## **Detection Range:**

0.312-20ng/mL

### **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

## **Tested Application:**

**ELISA** 

## **SwissProt:**

D4A3X8


# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated





with an antibody specific to Interleukin 7 Receptor (IL7R). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Interleukin 7 Receptor (IL7R). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Interleukin 7 Receptor (IL7R), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of Interleukin 7 Receptor (IL7R) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!