

# ELISA Kit for Nicotinamide Adenine Dinucleotide Phosphate Oxidase 4 (NOX4)

Catalog No: tcue1917



## Available Sizes

---

**Size:** 96T



## Specifications

---

**Research Area:**

Signal transduction;Enzyme & Kinase;Metabolic pathway;Tumor immunity;

---

**Species Reactivity:**

Homo sapiens (Human)

---

**Sample Type:**

Tissue homogenates, cell lysates and other biological fluids

---

**Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.057ng/mL

---

**Detection Range:**

0.156-10ng/mL

---

**Assay Time:**

3h

---

**Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

---

**Tested Application:**

ELISA

---


**SwissProt:**

Q9NPH5

---

## Test Principle

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Nicotinamide Adenine Dinucleotide Phosphate Oxidase 4 (NOX4). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Nicotinamide Adenine Dinucleotide Phosphate Oxidase 4 (NOX4). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Nicotinamide Adenine Dinucleotide Phosphate Oxidase 4 (NOX4), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450\text{nm} \pm 10\text{nm}$ . The concentration of Nicotinamide Adenine Dinucleotide Phosphate Oxidase 4 (NOX4) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!