



# **ELISA Kit for Glutathione S Transferase Omega 1** (GSTo1)

Catalog No: tcue2370



## **Available Sizes**

Size: 96T



# **Specifications**

#### **Research Area:**

Enzyme & Kinase; Metabolic pathway; Tumor immunity; Hepatology;

#### **Species Reactivity:**

Homo sapiens (Human)

#### **Sample Type:**

Serum, plasma and other biological fluids.

# **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.056ng/mL

# **Detection Range:**

0.156-10ng/mL

## **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

## **Tested Application:**

ELISA

#### **SwissProt:**


P78417





# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Glutathione S Transferase Omega 1 (GSTo1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Glutathione S Transferase Omega 1 (GSTo1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Glutathione S Transferase Omega 1 (GSTo1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Glutathione S Transferase Omega 1 (GSTo1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!