



# **ELISA Kit for Chemokine C-X-C-Motif Receptor 4** (CXCR4)

Catalog No: tcue2744



## **Available Sizes**

Size: 96T



# **Specifications**

#### **Research Area:**

Signal transduction; CD & Adhesion molecule; Tumor immunity; Infection immunity;

## **Species Reactivity:**

Sus scrofa; Porcine (Pig)

#### **Sample Type:**

Tissue homogenates and other biological fluids.

## **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.059ng/mL

#### **Detection Range:**

0.156-10ng/mL

## **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

## **Tested Application:**

ELISA

#### **SwissProt:**


Q764M9





# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Chemokine C-X-C-Motif Receptor 4 (CXCR4). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Chemokine C-X-C-Motif Receptor 4 (CXCR4). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Chemokine C-X-C-Motif Receptor 4 (CXCR4), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Chemokine C-X-C-Motif Receptor 4 (CXCR4) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!