



# **ELISA Kit for Pregnane X Receptor (PXR)**

Catalog No: tcue2788

4				
S	i	Z	e	•

### **Available Sizes**

96T



# **Specifications**

#### Research Area:

Signal transduction;

# **Species Reactivity:**

Homo sapiens (Human)

# **Sample Type:**

Tissue homogenates, cell lysates and other biological fluids

# **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.126ng/mL

# **Detection Range:**

0.312-20ng/mL

# **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

# **Tested Application:**

**ELISA** 

#### **SwissProt:**

075469

# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated





with an antibody specific to Pregnane X Receptor (PXR). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Pregnane X Receptor (PXR). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Pregnane X Receptor (PXR), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of  $450 \text{nm} \pm 10 \text{nm}$ . The concentration of Pregnane X Receptor (PXR) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!