



# **ELISA Kit for Fatty Acid Desaturase 2 (FADS2)**

**Catalog No: tcue3023** 

且	Available Sizes
Size:	96T

# **Specifications**

# **Research Area:** Enzyme & Kinase;

Liizyiiic & Kiiidse,

# **Species Reactivity:**

Rattus norvegicus (Rat)

#### **Sample Type:**

serum, plasma, tissue homogenates and other biological fluids

#### **Sensitivity:**

The minimum detectable dose of this kit is typically less than 0.64ng/mL

### **Detection Range:**

1.56-100ng/mL

#### **Assay Time:**

3h

#### **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

## **Tested Application:**

**ELISA** 

#### **SwissProt:**

Q9Z122


# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated





with an antibody specific to Fatty Acid Desaturase 2 (FADS2). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Fatty Acid Desaturase 2 (FADS2). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Fatty Acid Desaturase 2 (FADS2), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm ± 10nm. The concentration of Fatty Acid Desaturase 2 (FADS2) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!