

# ELISA Kit for Myxovirus Resistance 1 (MX1) Catalog No: tcue3467

**Available Sizes** 

# **Size:** 96T

Specifications

# **Research Area:**

Infection immunity;

Species Reactivity:

Homo sapiens (Human)

### Sample Type:

tissue homogenates, cell lysates, cell culture supernates and other biological fluids

# Sensitivity:

The minimum detectable dose of this kit is typically less than 0.65ng/mL

#### **Detection Range:**

1.56-100ng/mL

# **Assay Time:**

3h

# **Detection Method:**

Enzyme-linked immunosorbent assay for Antigen Detection.

# **Tested Application:**

ELISA

### SwissProt:


P20591

# **Test Principle**

The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated



with an antibody specific to Myxovirus Resistance 1 (MX1). Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody specific to Myxovirus Resistance 1 (MX1). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Myxovirus Resistance 1 (MX1), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of Myxovirus Resistance 1 (MX1) in the samples is then determined by comparing the O.D. of the samples to the standard curve.



All products are for RESEARCH USE ONLY. Not for diagnostic & therapeutic purposes!

Copyright 2021 Taiclone Biotech Corp.